

# Jandel Engineering Limited



## Jandel reference sample

*Jandel Engineering Limited* supplies an indium tin oxide (ITO) coated glass sample that can be used as a reference sample to assess whether a four point probing system is correctly set up and offering reasonable readings.

The sample is not NIST traceable but assigned values are carefully assessed based on measurements made by an employee with over twenty years of four point probe measurement experience.

The dimensions of the samples are 50mm square with the assigned value applying to the 25mm square central area of the sample. The assigned values are between 12 and 15 ohms-per-square and printed on the sample as shown above.

*If you require any further information on the Jandel Reference sample please do not hesitate to contact us using the details below*